



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005  
& ANSI/NCSL Z540-1-1994

TEAM TORQUE, INC.  
1231 Park Avenue  
Bismarck, ND 58504  
Jim Mueller Phone: 701 223 4552

CALIBRATION

Valid To: September 30, 2012

Certificate Number: 2472.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations<sup>1</sup>:

I. Electrical – DC & Low Frequency

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
DC Voltage – Measure	(1.0 to 100) mV 101 mV to 1 V (1.001 to 10) V (10.001 to 30) V	0.036 mV 0.064 mV 0.46 mV 5.9 mV	6.5 digit digital multimeter

## II. Mechanical

Parameter/Equipment	Range	CMC <sup>2,3</sup> (±)	Comments
Torque –  Wrenches	3.5 in·oz to 50 in·lb (51 to 150) in·lb (151 to 400) in·lb (401 to 1000) in·lb (25 to 250) ft·lb (251 to 600) ft·lb (601 to 1000) ft·lb (1001 to 2000) ft·lb	1.3 % 1.1 % 0.80 % 0.68 % 1.2 % 0.89 % 1.1 % 1.2 %	Torque tester
Watches	0.5 in·oz to 20 in·lb	0.79 %	
Screwdrivers	2.0 in·oz to 50 in·lb (51 to 400) in·lb	1.3 % 0.80 %	
Torque Transducers	0.5 in·oz to 500 in·lb (25 to 250) ft·lb (251 to 600) ft·lb (601 to 2000) ft·lb	0.12 % 0.24 % 0.39 % 0.56 %	Dead weights

<sup>1</sup> This laboratory offers commercial calibration service.

<sup>2</sup> Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

<sup>3</sup> In the statement of CMC, all percentages represent percent of range.





World Class Accreditation

The American Association for Laboratory Accreditation

# Accredited Laboratory

A2LA has accredited

**TEAM TORQUE, INC.**

*Bismarck, ND*

for technical competence in the field of

**Calibration**

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005 *General Requirements for the Competence of Testing and Calibration Laboratories*. This laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and any additional program requirements in the field of calibration. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (*refer to joint ISO-ILAC-IAF Communiqué dated 8 January 2009*).

Presented this 19<sup>th</sup> day of October 2010.



  
\_\_\_\_\_  
Peter Meyer

President & CEO  
For the Accreditation Council  
Certificate Number 2472.01  
Valid to September 30, 2012

*For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.*